



COMPRESSED GAS ASSOCIATION, INC.

SAFETY ALERT

SA-14—2002

4221 WALNEY ROAD, 5TH FLOOR
CHANTILLY, VA 20151
703-788-2700
cga@cganet.com

HAZARDS OF COMPRESSED GAS CYLINDERS IN THE MAGNETIC RESONANCE IMAGING (MRI) ENVIRONMENT

Compressed gas cylinders, cylinder valves, cylinder regulators, other assemblies included with cylinders, or any other equipment used to supply compressed gases shall not be present in the same room as an MRI device unless proven non magnetic and/or tested and deemed suitable for this environment.

Serious personal injury and property damage has resulted from the use of steel gas cylinders in close proximity to MRI devices. One incident resulted in the death of a six-year-old boy when a steel cylinder was brought into a room containing an operating MRI unit. The magnetic forces generated by the MRI drew the cylinder across the room, striking the boy in the head.

This is an area of serious concern to the Compressed Gas Association (CGA) and the focus of this safety alert is to notify personnel who handle, use, transport, or store compressed gas cylinders and equipment to this hazard.

If there are any questions as to whether a piece of equipment is suitable for use in the MRI environment, the equipment manufacturer or supplier should be contacted.

PLEASE NOTE:

The information contained in this document was obtained from sources believed to be reliable and is based on technical information and experience currently available from members of the Compressed Gas Association, Inc. and others. However, the Association or its members, jointly or severally, make no guarantee of the results and assume no liability or responsibility in connection with the information or suggestions herein contained. Moreover, it should not be assumed that every acceptable commodity grade, test or safety procedure or method, precaution, equipment or device is contained within, or that abnormal or unusual circumstances may not warrant or suggest further requirements or additional procedure.

This document is subject to periodic review, and users are cautioned to obtain the latest edition. The Association invites comments and suggestions for consideration. In connection with such review, any such comments or suggestions will be fully reviewed by the Association after giving the party, upon request, a reasonable opportunity to be heard. Proposed changes may be submitted via the Internet at our web site, www.cganet.com.

This document should not be confused with Federal, state, provincial, or municipal specifications or regulations; insurance requirements; or national safety codes. While the Association recommends reference to or use of this document by government agencies and others, this document is purely voluntary and not binding.

A listing of all publications, audiovisual programs, safety and technical bulletins, and safety posters is available via the Internet at our website at www.cganet.com. For more information contact CGA at Phone: 703-788-2700, ext. 799. E-mail: customerservice@cganet.com.

FIRST EDITION: 2002

© 2002 BY THE COMPRESSED GAS ASSOCIATION, INC. ALL RIGHTS RESERVED.
4221 WALNEY ROAD, 5TH FLOOR, CHANTILLY, VA 20151
Printed in U.S.A.